

Deep South Pickles, LLC
pH meter calibration log

% slope spec: 100% +/- 3%

<u>Date</u>	<u>Time</u>	<u>pH meter slope %</u>	<u>Comments</u>	<u>Initials</u>
5/1/2013	10:45	98	GOOD	ERE
5/1/2013	14:40	99	GOOD	ERE
5/7/2013	10:25	102	GOOD	ERE
5/7/2013	13:20	99	GOOD	ERE
5/10/2013	9:45	101	GOOD	RBG
5/10/2013	14:35	101	GOOD	RBG
5/16/2013	10:08	98	GOOD	RBG
5/16/2013	14:50	99	GOOD	RBG

Records reviewed by: _____

Date: _____